



PATENT

for
12-3-02

IN THE UNITED STATES PATENT AND TRADEMARK OFFICE

Applicant: Christopherson et al.

Docket: P-9126

Serial No.: 09/776,265

Group Art: 3736

Filed: February 2, 2001

Examiner: David J. McCrosky

Title: INFORMATION REMOTE MONITOR (IRM) MEDICAL DEVICE

RECEIVED

NOV 27 2002

TECHNOLOGY CENTER R3700

RESPONSE TO ADVISORY OFFICE ACTION

DATED OCTOBER 22, 2002

Honorable Commissioner For Patents and Trademarks
Washington, D.C. 20231

Sir:

Reconsideration of the above-identified application is respectfully requested.

REMARKS

The rejection of the claims as being obvious from Halperin (U.S. Patent No. 5,810,735) in view of Krichen (U.S. Patent No. 6,250,309) has been maintained in the Advisory Action. The basis for the rejection is that Halperin discloses an external barometric pressure sensor used in combination with an absolute pressure measurement obtained by an implanted cardiac pressure sensor to produce corrected cardiac pressure data that is telemetered to a programmer. Halperin, however, fails to disclose a communication link between the programmer and a remote data center over which corrected cardiac pressure data is transferred. Krichen is relied upon as teaching